Pulser – Programmable TTL Pulse Train Generator for Optogenetics Light Sources

Introduction

The Pulser is a programmable TTL pulse train generator for pulsing LEDs, lasers and shutters used in Optogenetics activation in neurophysiology and behavioral research. The Pulser device comes with user friendly Windows[®] software that enables easy visual setup of numerous pulse train configurations. The Pulser TTL-Out output can work smoothly to trigger Prizmatix LED light sources or lasers to produce light pulses for such application as opsin activation. The Trigger-In input enables synchronization of



Optogenetics activation with various experimental events. The Trigger-Out output (PulserPlus only) enables further activation of additional devices after pulse train was completed.

Key Features


- Single or Bi-phasic pulse trains
- Trigger-In input enables synchronization with various experimental events.
- Trigger-Out output (PulserPlus only) enables activation of other devices after completion of pulse train sequence.
- Operation modes:
 - Mode (0): Execute pulse sequence after START
 - Mode (1): Execute pulse sequence once after trigger(s)
 - Mode (2): Execute pulse sequence each time after trigger(s)
 - Mode (3): Execute pulse sequence after trigger HIGH, then stop when LOW
 - Mode (4): Execute pulse sequence after trigger LOW, then stop when HIGH
- Pulser settings can be stored into a file to enable fast switch between various protocols.
- Log-File can be created to record experiment activities.
- No external power supply required (uses USB power for operation)
- Stand-alone operation mode enables Pulser to operate without computer. The power is supplied by USB power adaptor (PulserPlus only)
- Control up to 8 Prizmatix Optogenetics-LEDs with a single Pulser device using BNC-T connectors (PulserPlus only)
- Independent time base by internal microcontroller
- Easily control from Noldus EthoVision XT, LabVIEW, MATLAB or any software capable of sending and receiving simple ASCII commands over serial COM port (API provided).

Main Office	European Sales Office	North America Sales Office		
Phone: +972-72-2500097	Phone: +44-(0)77-9172-9592	Phone: +1-(248)-436-8085		
Fax: +972-72-2500096	Fax: +44-(0)20-7681-2977	Fax: +1-(248)-281-5236		
sales@prizmatix.com	sales.europe@prizmatix.com	sales.usa@prizmatix.com		
P.O.	B. 244 Givat-Shmuel 54101,	Israel		

Setups


Possible Pulser setups:

- (A) PC Control with direct connection to LED.
- (B) PC Control with external trigger device.



PulserPlus can be used in two additional setups:

- (C) PC Control with trigger in and trigger out to send pulses to additional triggered device.
- (D) SA -Stand Alone mode, triggered or manually activated, no PC needed during operation.



Main Office	European Sales Office	North America Sales Office		
Phone: +972-72-2500097	Phone: +44-(0)77-9172-9592	Phone: +1-(248)-436-8085		
Fax: +972-72-2500096	Fax: +44-(0)20-7681-2977	Fax: +1-(248)-281-5236		
sales@prizmatix.com	sales.europe@prizmatix.com	sales.usa@prizmatix.com		
P.O.	B. 244 Givat-Shmuel 54101,	Israel		


Software

The software enables easy visual design of the various pulse conditions. The pictures below show the Graphical User Interface (GUI). The upper right side of the GUI shows an example waveform with the legend of user definable parameters. The bottom pane shows the actual TTL waveform programmed with the user settings on the left side.

File Options Help										
Group Settings Train Settings	Pulser Plus Settings	-	TI P10	P11 .	P2D	P21 1	I.			П
Total # of Trains	10		Н					ſ		
TI - Train Interval [ms]	0		_ l				11			-
P1D - Pulse 1 Duration [ms]			-	Trai	in ∦1			Train	#2	
P1I - Pulse 1 Interval [ms] P2D - Pulse 2 Duration [ms]	200	-		_		_		_		
P2I - Pulse 2 Interval [ms]										
Start Sequence	STOP							+		
EXIT		ò	100	200	300	400	500	600	700	800
Select COM port: % COM28 Select Operation Mode: () (0) Exec	cute pulse sequence aft	ter START	_	_	1	ime (mse		er Plus is	connecte	d

Train Settings Tab


The Train Settings tab allows setting of single or bi-phasic pulse timing and pulse repetition creating multiple pulse trains.



Group Settings Tab

The Group Settings tab allows creation of multiple Groups of Trains, comprising the sequence. The TrigIn parameter defines the number of triggers needed to start sequence if appropriate Operation Mode is chosen.

Main Office	European Sales Office	North America Sales Office		
Phone: +972-72-2500097	Phone: +44-(0)77-9172-9592	Phone: +1-(248)-436-8085		
Fax: +972-72-2500096	Fax: +44-(0)20-7681-2977	Fax: +1-(248)-281-5236		
sales@prizmatix.com	sales.europe@prizmatix.com	sales.usa@prizmatix.com		
P.O.E	3. 244 Givat-Shmuel 54101,	Israel		



PulserPlus Settings Tab

The Pulser-Plus is an enhanced TTL Pulser that adds stand-alone operation and TTL Trigger-Out.

The PulserPlus tab in GUI allows the user to define all the trigger-out parameters as well as switch between Stand-Alone programming mode and PC Run mode.

The Operation mode selector at the bottom of GUI provides four operation modes (see key features above).


Pulser vis. PulserPlus Comparison

Feature	Pulser	PulserPlus		
Independent time base by				
internal microcontroller	•	•		
Sequence Settings	♦	•		
Train Settings	♦	•		
Group Settings	♦	•		
Trigger In Settings	♦	•		
Trigger Out Setting	-	•		
Stand-Alone Operation	-	•		
Settings Files	♦	•		
Log File	♦	•		

Main Office	European Sales Office	North America Sales Office		
Phone: +972-72-2500097	Phone: +44-(0)77-9172-9592	Phone: +1-(248)-436-8085		
Fax: +972-72-2500096	Fax: +44-(0)20-7681-2977	Fax: +1-(248)-281-5236		
sales@prizmatix.com	sales.europe@prizmatix.com	sales.usa@prizmatix.com		
P.O.	B. 244 Givat-Shmuel 54101,	Israel		

Specifications

USB connector		Standard USB Type B
Signal connectors		BNC
Output voltage level		TTL
Minimum pulse width	msec	1



Dimensions in mm

Ordering Information

_

Item Description	P/N	
Programmable pulse train generator for use with Optogenetics-LED		
or other light sources.		
Connectors: USB, TTL Out, Trigger In	Dulaar	
Included:	Pulser	
1 x USB Type A cable		
1 x BNC-BNC cable		
Programmable pulse train generator for use with Optogenetics-LED		
or other light sources.		
Connectors: USB, TTL Out, Trigger In, Trigger Out		
Included:	PulserPlus	
1 x USB Type A cable	r uisei r ius	
1 x BNC-BNC cable		
Not included: Power adaptor for Stand-Alone operation. Any USB		
power adaptor can be used		

Main Office	European Sales Office	North America Sales Office		
Phone: +972-72-2500097	Phone: +44-(0)77-9172-9592	Phone: +1-(248)-436-8085		
Fax: +972-72-2500096	Fax: +44-(0)20-7681-2977	Fax: +1-(248)-281-5236		
sales@prizmatix.com	sales.europe@prizmatix.com	sales.usa@prizmatix.com		
P.O.	B. 244 Givat-Shmuel 54101,	Israel		